

MAY - 4 2005

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) number is: K050589.

1. Submitted by:	Sysmex America, Inc. One Nelson C. White Parkway Mundelein, IL 60060 Phone: (847) 996-4675; FAX: (847) 996-4655 Contact person: Nina Gamperling Date prepared: March 7, 2005
2. Name of Device:	<u>Trade or proprietary name:</u> RET-He parameter on the Sysmex [®] XE-2100, Automated Hematology Analyzer. <u>Common name:</u> RET-He <u>Classification name:</u> RET-He parameter on the Automated Differential Cell Counter, Sysmex [®] XE-2100 (21 CFR 864.5220)
3. Predicate Device:	The RET-He parameter on the Sysmex [®] XE-2100, Automated Hematology Analyzer, is substantially equivalent to the CHr parameter on the Bayer Advia 120 Hematology System.
4. Device Description:	The XE-2100 is an automated hematology analyzer previously cleared by the FDA. The RET-He parameter determines the hemoglobin of the reticulocytes. (Note: XE-Pro and RET Master are required to obtain results described.)
5. Intended Use:	The RET-He parameter on the Sysmex [®] XE-2100, Automated Hematology Analyzer, determines the hemoglobin of reticulocytes for <i>in vitro</i> diagnostic use in clinical laboratories.
6. Substantial equivalence-similarities and differences	The following table compares the RET-He parameter with predicate method.

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS (continued)

Comparison Table to Predicate Method

	CHr parameter on the Bayer Advia 120	RET-He parameter on the Sysmex XE-2100
	Predicate K971998	New method
Intended Use	The Advia 120 Hematology System is a quantitative, automated hematology analyzer that provides a leukocyte differential count and reticulocyte analysis for <i>in vitro</i> diagnostic use in clinical laboratories.	The RET-He parameter on the Sysmex® XE-2100, Automated Hematology Analyzer, determines the hemoglobin of reticulocytes for <i>in vitro</i> diagnostic use in clinical laboratories.
Methodology	The reticulocyte parameters are derived through a combination of laser light scatter and absorption of a nucleic acid dye.	The reticulocyte parameters are derived using the reticulocyte forward scattered light signals from the reticulocyte measurement channel and a proprietary Sysmex calculation equation.
Type of Anticoagulant	EDTA	EDTA
Specimen Type	Peripheral blood	Peripheral blood
Accuracy	Method of CHr, mean hemoglobin content of reticulocytes, has been established as the predicate method.	Comparison to CHr showed excellent correlation.

7. Clinical Performance Data:

Studies were performed to evaluate the equivalency of the RET-He parameter to the predicate method. Results indicated equivalent performance.

8. Conclusions:

The performance data demonstrated substantial equivalence.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

MAY - 4 2005

Ms. Nina M. Gamperling, MBA, MT(ASCP), RAC
Manager, Regulatory Affairs
Sysmex America, Inc.
One Nelson C. White Parkway
Mundelein, IL 60060

Re: k050589
Trade/Device Name: RET-He parameter on the Sysmex® XE-2100™, Automated Hematology Analyzer
Regulation Number: 21 CFR 864.5220
Regulation Name: Automated differential cell counter
Regulatory Class: Class II
Product Code: GKZ
Dated: March 7, 2005
Received: March 9, 2005

Dear Ms. Gamperling:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

Page 2 –

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0484. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>

Sincerely yours,

A handwritten signature in black ink, reading "Robert L. Becker, Jr." with a stylized flourish at the end.

Robert L. Becker, Jr., MD, Ph.D
Director
Division of Immunology and Hematology
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K050589

Device Name: RET-He parameter on the Sysmex® XE-2100™, Automated Hematology Analyzer

Indications For Use:

The RET-He parameter on the Sysmex® XE-2100, Automated Hematology Analyzer, determines the hemoglobin of reticulocytes for *in vitro* diagnostic use in clinical laboratories.

(PLEASE DO NOT WRITE BELOW THIS LINE- CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CHRD, Office of Device Evaluation (ODE)

Prescription Use ✓

OR

Over-The-Counter Use _____

Josephine Bantala
Division Sign Off

Office of In Vitro Diagnostic Device
Evaluation and Safety